

**GOVERNMENT OF ASSAM
DIRECTORATE OF MEDICAL EDUCATION, ASSAM.
HOUSEFED COMPLEX, DISPUR, GUWAHATI-6.**

FAX /PHONE NO:0361-2261629

No. DME/PET-CT/147/2007/

Dated 17-11-2008.

**INVITATION FOR BID (IFB)
GLOBAL COMPETITIVE BIDDING
CORRIGENDUM**

Following corrigendum is issued in connection with the IFB No. DME/PET-CT/147/2007/12216 dated 26-08-2008.

- A. Page 40 : Add in point iv “However this price shall not be considered for evaluation of price bid”
- B. Page 40 1.1 (i) Gantry and Detector: The PET Scanner should employ non-hygroscopic scintillation material add “LSO/LYSO with latest technology “ for detecting 511 KeV gamma photons in coincidence. is to be incorporated and delete the sentence “crystal material with decay rate below 60 nano second and should be mentioned in the offer”
- C. Page 42 (4 iii)] Replace sentence System efficiency/sensitivity must exceed 4cps/KBq at 350 KeV” by “system efficiency/ sensitivity must be specified by the vendor”
- D. Page 41(2.i.)] Replace Multi slice CT having capability of 32 slices or more per revolution by ” Multi slice CT having capability of 64 slices or more per revolution”

Part II (Medical Cyclotrons)

- A. Page 46 (vi) The word “Bunker” should be incorporated after self shielded with a slash.
- B. Page 47 Point (vii) the words “at least five “ should be omitted.
- C. Page 47 Point 2(i) the word “Bunker” should be added after self shielded with a slash.
- D. Page 48 Point 2 (iv) Replace the sentence “Beam exit ports-System should be capable of bombarding two targets simultaneously and each target port should have facility to mount at least 4 targets” by “Beam exit ports-System should be capable of bombarding two targets simultaneously and each target port should have facility to mount multiple target ports”.
- E. Page 50 point 7(ii) Replace the word “one” by “Two”
- F. Page 53 Part (III) point (iii): The figure “7000 sq feet” to be replaced by approximately “10000 sq ft”
- G. Page 53 Point (i) omit the word (DORMA make).
- H. Page 54 Point (V) omit the words “ of PGIMER,FM 200 gas flooding system should also be installed”.
- I. Page 56: Point(v).Scheduling: The number “12” should be replaced by “6”
- J. Page 57 Point (xi): The figure “7000 sq feet” to be replaced by approximately “10000 sq ft”.

All bidders must submit an undertaking mentioning that they have not supplied similar equipment in price less than the quoted price here to any other Institution / Organization in India in last 2 years.

The last date of submission of the Bid is extended up-to 15th December 2008 at 1 P.M and will be opened on the same date at 2 P.M. Other terms and conditions of the IFB No. DME/PET-CT/147/2007/12216 dated 26-08-2008 remains unchanged.

Sd./-Pro. U.C. Sarma,
Director of Medical Education, Assam.

Memo No. DME/147/2007/

Dated: 17-11-2008

Copy forwarded for information and necessary action to:

1. The Secretary to the Govt. of Assam, Health & F.W. (B) Dept., Dispur, Guwahati-06 for favour of kind information.
2. The Director of Information & Public Relations, Assam, Dispur, Guwahati-06. He is requested to publish the above corrigendum in those news papers where our Invitation for Bid (IFB) Global Competitive Bidding IFB No.DME/PET-CT/147/2007/12216, dated 26-08-2008 had been earlier published. Enclosed: 6 (six) copies.
3. The Mission Director, NRHM, Assam. He is requested kindly to upload the corrigendum in the NRHM, Assam Website. Soft copy enclosed.
4. The State Informatics Officer, Block F, Assam Secretariat, Dispur, Guwahati-06. He is requested to upload the corrigendum in the Assam Govt. Website. Soft copy enclosed.
5. The P.S. to the Hon'ble Minister, Health & F.W. Assam for kind appraisal of the Hon'ble Minister.
6. Notice Board of this Directorate.

Director of Medical Education, Assam.